



## Eagle Eye Power Solutions DM-Series Site Survey Form

### Objective

The purpose of this form is to gather necessary information regarding the site and application for our DM-Series Inline Density Meters. This information is required before production of the order is started. Please fill out all information possible so we can ensure all of your needs are addressed. Filling out this form does not obligate you to purchase the device. It is used only to assist in recommending the appropriate model for your application.

**\*\*Please do not hesitate to contact Eagle Eye Power Solutions for support with filling out this form.**

### Customer Information

<b>Name:</b>	<input type="text"/>	<b>Company Name:</b>	<input type="text"/>
<b>Work Number:</b>	<input type="text"/>	<b>Mobile Number:</b>	<input type="text"/>
<b>Email:</b>	<input type="text"/>		

### Site Information

<b>Environment:</b>	<b>Tank or Pipeline Mounting: *(2)</b>	<b>Need Bypass Line:</b>
<input type="checkbox"/> Industrial *(1)	<input type="checkbox"/> Tank	<input type="checkbox"/> Yes
<input type="checkbox"/> Sanitary	Tank Gap <input type="text"/>	<input type="checkbox"/> No
<input type="checkbox"/> Other <input type="text"/>	<input type="checkbox"/> Pipeline	
	Pipeline Size <input type="text"/>	

### Solution Information

<b>Description:</b>	<input type="text"/>	<b>Density:</b>	<input type="text"/>
<b>Temperature: *(3)</b>	<input type="text"/>	<b>Pressure:</b>	<input type="text"/>
<b>Flow Speed:</b>	<input type="text"/>		

### Model Information/Features

<b>Measuring Range: *(4)</b>	<b>Power Supply:</b>	<b>Communication Output:</b>
<input type="checkbox"/> 0 – 1 g/cm <sup>3</sup>	<input type="checkbox"/> 15 – 30 VDC	<input type="checkbox"/> 4 – 20 mA (HART Protocol)
<input type="checkbox"/> 0 – 2 g/cm <sup>3</sup>	<input type="checkbox"/> Other <input type="text"/>	<input type="checkbox"/> None
<input type="checkbox"/> 0 – 3 g/cm <sup>3</sup>		

**Length Between Sensors: \*(5)**

- 200 mm (8 in)
- 500 mm (20 in)
- 1000 mm (39 in)
- Other

**Installation: \*(6)**

- A-Style (Vertically Top-Mounted)
- B-Style (Vertically Side-Mounted)
- C-Style (Vertically Side-Flex-Mounted)

**Flange Size: \*(7)**

- DN65 (7 in OD)
- DN80 (7.5 in OD)
- DN100 (9 in OD)
- Not Sure

**Pictures Checklist**

Please provide (attach) and check off the images listed below.

Please indicate direction of flow on at least one image or in a drawing.

- Tank/Pipe Setup (Front)       Tank/Pipe Setup (Side)

**Comments**

Please provide any additional comments in the field below.

**\*Notes**

- (1) Industrial models not fit for use in food & beverage applications.
- (2) If installing to tank, please list (mm or in.) the distance between the mounting position (top of tank) and water level. If installing to pipeline, please list the pipeline size.
- (3) Sample temperature range based on model. Normal models range 0-100°C (32-212°F). Vertically Side- Flex-Mounted models range 0-150°C (32-302°F).
- (4) Accuracy varies depending on measuring range. Closer range (0-1 g/cm<sup>3</sup>) has higher accuracy at ±0.001 g/cm<sup>3</sup>. Medium range accuracy (0-2 g/cm<sup>3</sup>) is ±0.002-0.003 g/cm<sup>3</sup>. Large range accuracy (0-3 g/cm<sup>3</sup>) is ±0.002 – 0.005 g/ cm<sup>3</sup>. Temperature accuracy is not affected.
- (5) The spacing between sensors must be at least 200 mm (8 in) and the diameter 50 mm (2 in). If selecting the Vertically Side-Flex-Mounted meter, please measure the size of the flanges and the length of the pipes inserted into the liquid. Sensors must be fully submerged in solution. If needed, the distance between the main flange and first (top) sensor can be adjusted.
- (6) See the [DM-Series Inline Density Meters](#) for more details on various installation shapes.
- (7) Flange size spec based on ASME standard sizes.